

BACTERIOLOGICAL TEST REPORT



Client: Scientific Sanitation Solutions
#Address: Scientific Sanitation Solutions
Unit 7
67 Regency Drive
Route 21 Corporate Park
Irene
RSA

Job Number: M/2148/86249
Report No.: 104745.1
#Date Tak'n: Tue, 16th Feb 2021 10:00
Date Rec'd: Tue, 16th Feb 2021

Stansted Laboratories LTD
Unit 9-10, Riverside Industrial Estate
27 Thames Road
Barking, Essex
IG11 0ND
Tel: +44 (0)20 8594 5104
Fax: +44 (0)20 8591 8762
sales@stanstedlabs.co.uk
www.stanstedlabs.co.uk

#Site: Scientific Sanitation Solutions Unit 7
67 Regency Drive Route 21
Corporate Park Irene RSA

#FAO: Stuart Danby, Victor Adendorff

#Order No:

Directors: B. Patel, K.Patel

Items Marked \square are not included in the UKAS Schedule

Sample ID	#Description & Temperatures	Analysis	Result	Tested	Notes / Species
SL388575 <i>Miscellaneous</i>	SANI-99 Powder Init Temp \square Temp > 1 Min \square Temp > 2 Min \square	Legionella \square	N.I.	17/02/2021	None isolated from 500ml after 10 days incubation.
SL388576 <i>Miscellaneous</i>	Lab Legionella Control Init Temp \square Temp > 1 Min \square Temp > 2 Min \square	Legionella \square	2475 cfu/500ml	17/02/2021	Legionella pneumophila Serogroup 1

Method Codes
SLM/B380/M - Legionella spp

Legend & Footnotes
cfu = Colony Forming Units
mpn = Most Probable Number
TNTC = Too Numerous to Count
#Information supplied by the customer and can affect the validity of results.
Sampling is not included in our UKAS Schedule.
All Samples Analysed as received.

Test report authorised by :
B.Patel

Director
Date: 02/03/2021 15:40:00

Test Report checked by : Vasilena Petrova
Ivanova 02/03/2021 09:43:44

Disclaimer
Results relate to water samples tested and should not be reproduced except in full, without the written approval of the laboratory. Interpretations and opinions are not included in our UKAS schedule.
00:00 means time not declared.

----- END OF REPORT -----

Report No 104745.1 supersedes Report No. 104744.1 as the description has been amended.